PHYSIO CONTROL

LIFEPAK CR® Plus and LIFEPAK EXPRESS® DEFIBRILLATORS

User's Checklist

Year: _____

Unit Serial Number: _____

Location: _____

Month																											
Day																											
Initials																											
Instructions																											
Check defibrillator for damage or foreign substances.																											
OK is visible on readiness display. If not, refer to Troubleshooting Table.																											
Note Use By date of electrode pads:																											
Extra set of electrode pads stored with AED (optional). Note Use By date of electrode pads:																											
Infant/Child electrodes stored with AED (optional). Use By date:																											
Resuscitation Kit (e.g. AMBU® kit) containing: disposable gloves, face mask, trauma scissors and razor stored with AED.																											
Other resuscitation equipment stored with AED:																											

Troubleshooting Table

Readiness Display	Recommended Corrective Action
CHARGE-PAK™ indicator CHARGE-PAK battery charger and electrode pads need to be replaced OR CHARGE-PAK battery charger is not installed.	Replace CHARGE-PAK battery charger and QUIK-PAK™ electrode packet. NOTE: Do not install a CHARGE-PAK battery charger from another defibrillator.
ATTENTION indicator	Internal battery not fully charged. Verify CHARGE-PAK battery charger is installed. If CHARGE-PAK battery charger is installed use unit if needed. Replace CHARGE-PAK battery charger after each use of the defibrillator.
WRENCH indicator	Service required. Contact authorized service personnel only. In the USA, contact Physio-Control at 1.800.442.1142. Outside the USA, contact your local Physio-Control representative.

REMEMBER: Always attempt to use the defibrillator if needed in an emergency.
NOTE: If corrective action is needed leave box unchecked, complete troubleshooting log, and notify: _____.

Troubleshooting Log

Date/Time	Action Taken

Signature Log

Initials	Print Full Name

Initials	Print Full Name

©2014 Physio-Control Inc. Redmond, WA. All names herein are trademarks or registered trademarks of their respective owners.
GDR 3302426_C