



LUCAS[®] 2|3

Chest Compression System

GENUINE ACCESSORIES FROM PHYSIO-CONTROL

LUCAS 2|3 Chest Compression System





Thank you for choosing Physio-Control as your partner in helping save lives and improve patient care.

The LUCAS 3 chest compression system is shipped with one battery (dark grey), patient straps, stabilization strap, two suction cups, a hard shell carrying case and the Instructions for Use.

The LUCAS 2 chest compression system is shipped with one battery (light grey), patient straps, stabilization strap, three suction cups, a soft carrying case and the Instructions for Use.

In this catalog you will find the overview of accessories designed to meet your needs and optional power solutions. If you have any questions about accessories or disposables, please contact your Physio-Control representative.

LUCAS 2|3 Chest Compression System

Accessories for LUCAS 3 devices



LUCAS Carrying Case, Hard Shell

The LUCAS Hard Shell Case with Top Window for quick battery check and Rear Charge Port is the best way to protect your LUCAS device and carry your spare battery and accessories. Compatible with LUCAS 3 and LUCAS 2 devices.

11576-000081



LUCAS Back Plate, Slim

The slim, low profile back plate is easier to deploy, adjust, strap to transportation devices and clean. Compatible with LUCAS 3 and LUCAS 2 devices.

11576-000088



LUCAS Bumper integrated shaft seal, Black pair

The black rubber bumper accessory matches the colors of the LUCAS 3 device and is available to protect the hinge end caps.

11576-000091



Anti-slip tape, LUCAS Back Plate, Slim

The anti-slip tape is designed to increase the friction of the back plate and to reduce the risk of the device slipping when in contact with other hard surfaces (e.g., spine boards). Six strips in each pack.

11576-000089

11576-000090 (3-Pack)

Accessories for LUCAS 2 devices



LUCAS 2 Carrying Bag

LUCAS 2 device is stored and carried in a soft padded carrying bag complete with storage pouches for a spare battery and external power supply.

11576-000038



LUCAS 2 Back Plate

Replacement LUCAS 2 Back Plate is available in the event of damage or contamination of original.

21996-000044



LUCAS 2 Back Plate Anti-slip Tape

The LUCAS 2 Back Plate anti-slip tape is designed to increase the friction of the back plate and to reduce the risk of the device slipping when in contact with other hard surfaces (e.g., spine boards).

11576-000052

11576-000053 (3-Pack)



LUCAS 2 Rubber Bumper with Integrated Shaft Seal (Pair)

If the hinges protected by the rubber bumper on your device support legs match Fig 1, please order the Rubber Bumper with the integrated shaft seal (pair). The majority of LUCAS 2 devices require this bumper.

11576-000072

If you are still unsure which part you need, please contact customer support.




LUCAS 2 Rubber Bumper (Pair)

If your device has a plastic cap on the support leg hinges matching Fig 2, please order the Rubber Bumper without integrated shaft seal (pair). LUCAS 2 devices older than 2012 may require this bumper.

11576-000070

If you are still unsure which part you need, please contact customer support.



Product DVD

The LUCAS product DVD features an introduction on the physiology of CPR with Tim Phalen followed by a LUCAS 1 Inservice video and a LUCAS 2 Inservice video. A great DVD to have for new or refresher training.

21996-000043 (Inservice DVD)

Can also be found on physio-control.com/LUCAS

LUCAS 2|3 Chest Compression System

Shared Disposables, Batteries & Power Accessories

For use with LUCAS 3 or LUCAS 2 devices



LUCAS Disposable Suction Cup

LUCAS device has a disposable suction cup which can be easily changed after each use.

11576-000046 (3-Pack)
11576-000047 (12-Pack)



LUCAS Stabilization Strap

The cushion (new easier to clean material) and straps are used to provide further stabilization relative to the patient and to prevent downward movement of the LUCAS device during operation. Each package includes the neck strap and a pair of support leg straps.

21576-000074
21576-000075 (4-Pack)



LUCAS Patient Straps

The LUCAS patient straps are used to secure the patients arms to the support legs of the LUCAS device during transportation.

11576-000050 (Pair)
11576-000051 (Pair, 3-Pack)



LUCAS Battery

The LUCAS battery is a rechargeable Lithium-Polymer (LiPo) battery which typically last for 45 minutes of operation. You can charge the battery quickly, less than 2 hours in the device if connected to the external power supply or car power cable, and less than 4 hours in the desktop charger.

11576-000080 (Dark Grey-LUCAS 3 device)
11576-000039 (Light Grey-LUCAS 2 device)

Either battery will work in either device, color is the only difference.



LUCAS Battery Desk-Top Charger

The external battery desk-top charger is a stand-alone unit intended to charge one LUCAS Battery. It typically takes less than 4 hours to charge the battery. A mounting bracket is included for space-saving installation. New dark grey color.

11576-000060 (U.S., Canada)
11576-000061 (EU)
11576-000062 (UK)
11576-000063 (Australia)
11576-000068 (Japan)



LUCAS Aux Power Supply (S/N 3009 0181 and above)

The external power supply (100-240VAC, 50/60 Hz) can be connected to the LUCAS device and supports prolonged operation as well as charging of the LUCAS Battery while in the device.

11576-000071 (U.S., Canada)
11576-000056 (EU)
11576-000057 (UK)
11576-000058 (Japan)
11576-000059 (Australia)
11576-000067 (Switzerland)



LUCAS Car Cable

The car cable (12-24V DC) can be connected to the LUCAS 2 or LUCAS 3 devices. It supports prolonged operation as well as less than 2 hour charge time of the LUCAS 2 or LUCAS 3 Battery while in the device.

11576-000048



LUCAS PCI Back Plate

The carbon fiber LUCAS PCI Back Plate is intended specifically for use in the cath lab. It is fully radiotranslucent, with minimum shadows. The PCI Back Plate is compatible with LUCAS 2 and LUCAS 3 devices.


11576-000064

Instructions for Use and Training



LUCAS 3 Instructions for Use - (Software Version 3.0)

26500-003716 English (EN)	26500-003734 Turkish (TR)
26500-003717 International-English (INTL-EN)	26500-003735 Croatian (HR)
26500-003718 German (DE)	26500-003736 Chinese (ZH)
26500-003719 French (FR)	26500-003737 Slovakian (SK)
26500-003720 Dutch (NL)	26500-003738 Korean (KO)
26500-003721 Danish (DA)	26500-003739 Hebrew (HE)
26500-003722 Swedish (SV)	26500-003740 Bulgarian (BG)
26500-003723 Norwegian (NO)	26500-003741 Romanian (RO)
26500-003724 Finnish (FI)	26500-003742 Estonian (ET)
26500-003725 Italian (IT)	26500-003743 Latvian (LV)
26500-003726 Greek (EL)	26500-003744 Lithuanian (LT)
26500-003727 Spanish (ES)	26500-003745 Slovenian (SL)
26500-003728 Portuguese (PT)	26500-003746 Icelandic (IS)
26500-003729 Czech (CS)	26500-003747 Serbian (SR)
26500-003730 Polish (PL)	
26500-003731 Hungarian (HU)	
26500-003732 Russian (RU)	
26500-003733 Japanese (JA)	



LUCAS 2 Instructions for Use - (Software Version 2.2)

26500-003588 English (EN)	26500-003606 Turkish (TR)
26500-003589 International-English (INTL-EN)	26500-003607 Croatian (HR)
26500-003590 German (DE)	26500-003608 Chinese (ZH)
26500-003591 French (FR)	26500-003609 Slovakian (SK)
26500-003592 Dutch (NL)	26500-003610 Korean (KO)
26500-003602 Danish (DA)	26500-003648 Hebrew (HE)
26500-003593 Swedish (SV)	26500-003649 Bulgarian (BG)
26500-003594 Norwegian (NO)	26500-003650 Romanian (RO)
26500-003595 Finnish (FI)	26500-003651 Estonian (ET)
26500-003596 Italian (IT)	26500-003652 Latvian (LV)
26500-003597 Greek (EL)	26500-003653 Lithuanian (LT)
26500-003598 Spanish (ES)	26500-003654 Slovenian (SL)
26500-003599 Portuguese (PT)	
26500-003600 Czech (CS)	
26500-003601 Polish (PL)	
26500-003603 Hungarian (HU)	
26500-003604 Russian (RU)	
26500-003605 Japanese (JA)	

For version 2.1 please contact customer service.



LUCAS 3 or LUCAS 2 Training Device

The LUCAS 3 and LUCAS 2 Training Device is a non-clinical device intended for training purposes only. It should only be used on manikins.

LUCAS 3 99576-000042
LUCAS 2 99576-000020



For further information please contact your local Physio-Control representative or visit our website at www.physio-control.com



Physio-Control Headquarters
11811 Willows Road NE
Redmond, WA 98052 USA
Tel 425 867 4000
Fax 425 867 4121
www.physio-control.com

**Physio-Control Operations
Netherlands B.V.**
HQ Europe, Middle East & Africa
UP Office Building
Piet Heinkade 55,
1019 GM Amsterdam NL
Tel +31 (0)20 7070560
Fax +31 (0)20 3391194
www.physio-control.com

**Physio-Control
UK Sales Ltd**
11 Old Jewry, 7th Floor
London EC2R 8DU
United Kingdom

**Physio-Control
Australia Pty Ltd**
Suite 4.01
15 Orion Road
Lane Cove
NSW 2066
Australia
Toll Free Tel 1800 987 982
Toll Free Fax 1800 890 892