

Ositech Titan II Wireless Gateway

For use of this product in countries that allow higher Wi-Fi[®] channels than 1-11, please ensure the Wi-Fi Access Point being used is set to operate with Channels 1-11 only. Do not set “Auto Channel Selection” feature in the Access Point.

If an Access Point attempts to connect with a channel above Channel 11, the Wireless Gateway will fail to detect or connect to the Access Point.

For further information please contact your local Physio-Control representative.